



TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017

Blue = New Analyte

Magenta = Changes

Matrix	EPA Analyte Code	NELAC Analyte Code	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷
					a	b	c	d	
Microbiology				CFU/100 mL					CFU/100 mL
NPW	0233	2500	Total Coliform, MF ⁸	20 to 2400	Log transform; ±3 SD				2
NPW	0235	2530	Fecal Coliform, MF ⁸	20 to 2400	Log transform; ±3 SD				2
NPW		2525	E.coli, MF ⁸	20 to 2400	Log transform; ±3 SD				2
NPW		2520	Enterococci, MF ⁸	20 to 1000	Log transform; ±3 SD				2
				MPN/100 mL					MPN/100 mL
NPW	0234	2500	Total Coliform, MPN ⁹	20 to 2400	Log transform; ±3 SD				2
NPW	0236	2530	Fecal Coliform, MPN ⁹	20 to 2400	Log transform; ±3 SD				2
NPW		2525	E.coli, MPN ⁹	20 to 2400	Log transform; ±3 SD				2
NPW		2520	Enterococci, MPN ⁹	20 to 1000	Log transform; ±3 SD				2
Trace Metals				µg/L					µg/L
NPW	0001	1000	Aluminum	200 to 4000	0.9823	9.5889	0.0471	11.2110	144
NPW	0016	1005	Antimony	90 to 900	0.9864	-1.1174	0.0471	6.1230	57
NPW	0002	1010	Arsenic	90 to 900	0.9916	1.2647	0.0422	5.1741	64
NPW	0237	1015	Barium	100 to 2500	±15% fixed acceptance limit				85
NPW	0003	1020	Beryllium	50 to 500	±15% fixed acceptance limit				42
NPW		1025	Boron	800 to 2000	±15% fixed acceptance limit				680
NPW	0004	1030	Cadmium	100 to 1000	±15% fixed acceptance limit				85
NPW	0006	1040	Chromium, total	100 to 1000	±15% fixed acceptance limit				85
NPW	0238	1045	Chromium VI	90 to 900	0.9917	1.0232	0.0476	2.2011	71
NPW	0005	1050	Cobalt	100 to 1000	±15% fixed acceptance limit				85
NPW	0007	1055	Copper	100 to 1000	±15% fixed acceptance limit				85
NPW	0008	1070	Iron	200 to 4000	±15% fixed acceptance limit				170
NPW	0012	1075	Lead	100 to 1500	±15% fixed acceptance limit				85
NPW	0010	1090	Manganese	200 to 2000	±15% fixed acceptance limit				170
NPW	0009	1095	Mercury ^{10a}	3.0 to 30	±30% fixed acceptance limit				0.9
NPW	0074	1100	Molybdenum	60 to 600	0.9953	-0.1614	0.0372	2.5555	45
NPW	0011	1105	Nickel	200 to 2000	1.0012	1.5795	0.0368	3.8151	168
NPW	0013	1140	Selenium	100 to 1000	±15% fixed acceptance limit				85
NPW	0017	1150	Silver	100 to 1000	±15% fixed acceptance limit				85
NPW	0075	1160	Strontium	50 to 500	±15% fixed acceptance limit				42
NPW	0018	1165	Thallium	80 to 800	0.9932	-0.9634	0.0479	4.2361	54
NPW	0239	1175	Tin	200 to 2000	±30% fixed acceptance limit				140
NPW	0076	1180	Titanium	60 to 300	±15% fixed acceptance limit				51
NPW	0014	1185	Vanadium	50 to 2000	±15% fixed acceptance limit				42
NPW	0015	1190	Zinc	300 to 2000	±15% fixed acceptance limit				255



TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017

Blue = New Analyte Magenta = Changes

Matrix	EPA Analyte Code	NELAC Analyte Code	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷
					a	b	c	d	
			Demands ^{10b}	mg/L					mg/L
NPW	0038	1530	5-day BOD ^{10c}	18 to 230	0.6237	0.7022	0.0928	0.6636	4.9
NPW	0102	1555	Carbonaceous BOD ^{10c}	18 to 230	0.5648	0.6665	0.0965	0.8253	3.1
NPW	0036	1565	COD ^{10d}	30 to 250	0.9843	-0.3171	0.0432	3.0191	16
NPW	0037	2040	TOC ^{10e}	6.0 to 100	0.9926	0.1680	0.0473	0.3536	4.2
			Minerals	mg/L					mg/L
NPW	0027	1505	Alkalinity, total (CaCO ₃)	25 to 400	±20% at < 40; ±15% at ≥ 40 fixed acceptance limit				20
NPW		1540	Bromide	1.0 to 10	1.0098	-0.0533	0.0400	0.0912	0.56
NPW	0023	1035	Calcium	10 to 100	±15% fixed acceptance limit				8.5
NPW	0028	1575	Chloride	35 to 275	1.0005	0.0490	0.0376	0.3716	30
NPW	0029	1730	Fluoride	0.4 to 4	0.9748	0.0156	0.0487	0.0277	0.26
NPW		1550	Calcium hardness as CaCO ₃	25 to 250	±15% fixed acceptance limit				21
NPW	0022	1755	Hardness, total (CaCO ₃)	40 to 415	±15% fixed acceptance limit				34
NPW	0024	1085	Magnesium	4.0 to 40	±15% fixed acceptance limit				3.4
NPW	0026	1125	Potassium	4.0 to 40	±20% fixed acceptance limit				3.2
NPW	0025	1155	Sodium	10 to 100	±20% fixed acceptance limit				8.0
NPW	0020	1610	Spec. Cond. (25°C)	200 to 1200 µmhos/cm	±10% fixed acceptance limit				180 µmhos/cm
NPW	0030	2000	Sulfate	5.0 to 125	0.9880	-0.2130	0.0473	0.3309	3.0
NPW		2005	Sulfide	2.0 to 10	0.9657	-0.1271	0.1205	0.2816	0.20
NPW	0021	1955	Total Dissolved Solids at 180°C	140 to 800	1.0000	0.0000	0.0000	15.0000	95
NPW	0105	1950	Total Solids	140 to 800	1.0000	0.0000	0.0000	15.0000	95
			Nutrients	mg/L					mg/L
NPW	0031	1515	Ammonia as N	1.0 to 20	0.9923	0.0567	0.0583	0.0914	0.60
NPW	0032	1810	Nitrate as N	2.0 to 25	0.9975	-0.0005	0.0506	0.0642	1.50
NPW		1820	Nitrate-nitrite as N	2.5 to 25	0.9957	-0.0010	0.0509	0.0400	1.99
NPW		1840	Nitrite as N	0.4 to 4.0	1.0017	-0.0030	0.0377	0.0250	0.28
NPW	0033	1870	Orthophosphate as P	0.5 to 5.5	±15% fixed acceptance limit				0.42
NPW	0034	1795	Total Kjeldahl-Nitrogen ^{10f}	3.0 to 35	0.9701	0.2283	0.0680	0.1906	1.95
NPW	0035	1910	Total Phosphorus	0.5 to 10	0.9932	0.0084	0.0506	0.0254	0.35



TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017

Blue = New Analyte Magenta = Changes

Matrix	EPA Analyte Code	NELAC Analyte Code	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷
					a	b	c	d	
			Misc. Analytes	mg/L					mg/L
NPW		1500	Acidity, as CaCO ₃	650 to 1800	±10% fixed acceptance limit				585
NPW		1605	Color	10 to 75 PC units	0.9474	0.6098	0.0367	2.4407	1.7 PC units
NPW	0072	1960	Total Suspended Solids	20 to 100	0.9728	-0.6338	0.0300	1.5793	12
NPW	0019	1900	pH ^{10g}	5.0 to 10 units	± 0.2 units fixed acceptance limit				Not applicable
NPW	0071	1645	Total Cyanide ^{10h}	0.1 to 1	±35% fixed acceptance limit				0.065
NPW	0097	1905	Total Phenolics (4AAP) ¹⁰ⁱ	0.5 to 5	0.6408	0.0250	0.1038	0.0082	0.16
NPW	0098	1940	Total Residual Chlorine	0.5 to 3.0	0.9345	0.0392	0.0688	0.0073	0.38
NPW		1965	Settleable solids	5.0 to 50 mL/L	1.0436	-0.0108	0.0597	0.4546	2.9 mL/L
NPW		1990	Silica as SiO ₂	50 to 250	±25% fixed acceptance limit				38
NPW		2025	Surfactants - MBAS	0.2 to 1.0	1.0421	-0.0068	0.1326	0.0046	0.10
NPW		2055	Turbidity ^{10j}	2.0 to 30 NTU	1.0040	-0.0368	0.0475	0.1575	1.2 NTU
NPW		1970	Volatile solids, Total	100 to 500	0.9644	-4.7559	0.0182	14.9450	41
			Low Level Analytes ¹¹						
NPW		1095	Mercury ^{10a}	20 to 100 ng/L	0.9910	0.2064	0.0432	2.5774	9.7
NPW		1940	Total Residual Chlorine	50 to 250 µg/L	1.0000	0.0000	0.0000	20.0000	5.0



TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017

Blue = New Analyte Magenta = Changes

Matrix	EPA Analyte Code	NELAC Analyte Code	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷
					a	b	c	d	
			Volatile Aromatics ¹	µg/L					µg/L
NPW	0065	4375	Benzene	10 to 120		±30% fixed acceptance limit			7.0
NPW	0094	4610	1,2-Dichlorobenzene	10 to 120		±30% fixed acceptance limit			7.0
NPW	0096	4615	1,3-Dichlorobenzene	10 to 120		±30% fixed acceptance limit			7.0
NPW	0095	4620	1,4-Dichlorobenzene	10 to 120		±30% fixed acceptance limit			7.0
NPW	0066	4765	Ethylbenzene	10 to 120		±30% fixed acceptance limit			7.0
NPW	0222	5005	Naphthalene	15 to 150	0.8785	1.4343	0.1335	0.7561	6.3
NPW		5100	Styrene	20 to 120		±35% fixed acceptance limit			13.0
NPW	0067	5140	Toluene	10 to 120		±30% fixed acceptance limit			7.0
NPW	0092	5155	1,2,4-Trichlorobenzene	15 to 150	0.9160	-1.3028	0.1473	0.5100	4.3
NPW		5210	1,2,4-Trimethylbenzene	10 to 120		±35% fixed acceptance limit			6.5
NPW		5215	1,3,5-Trimethylbenzene	10 to 120		±35% fixed acceptance limit			6.5
NPW		5240	m/p-Xylenes	10 to 150		±40% fixed acceptance limit			6.0
NPW		5250	o-Xylene	10 to 150		±40% fixed acceptance limit			6.0
NPW	0242	5260	Xylenes, total ¹²	20 to 300		±40% fixed acceptance limit			12
			Volatile Ketones/Ethers ¹	µg/L					µg/L
NPW		4315	Acetone	20 to 200	0.8856	3.5838	0.2028	1.7474	3.9
NPW		4860	2-Hexanone	20 to 200	1.0054	-1.1748	0.1534	1.7764	4.4
NPW		4995	4-Methyl-2-pentanone (MIBK)	20 to 200	1.0022	-1.0337	0.0934	4.1819	2.0
NPW		5000	Methyl tert-butyl ether (MTBE)	15 to 150	1.0233	-0.3620	0.1112	0.3083	9.0



TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017

Blue = New Analyte Magenta = Changes

Matrix	EPA Analyte Code	NELAC Analyte Code	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷
					a	b	c	d	
			Volatile Halocarbons ¹	µg/L					µg/L
NPW	0060	4395	Bromodichloromethane	10 to 100		±40% fixed acceptance limit			6.0
NPW	0062	4400	Bromoform	10 to 100		±40% fixed acceptance limit			6.0
NPW	0243	4950	Bromomethane	20 to 120		± 60% fixed acceptance limit			8.0
NPW	0058	4455	Carbon tetrachloride	15 to 150	0.9577	0.0612	0.1269	0.3443	7.7
NPW	0064	4475	Chlorobenzene	10 to 120		±30% fixed acceptance limit			7.0
NPW	0244	4485	Chloroethane	20 to 120		± 60% fixed acceptance limit			8.0
NPW	0055	4505	Chloroform	10 to 100		±30% fixed acceptance limit			7.0
NPW	0245	4960	Chloromethane	20 to 120		± 60% fixed acceptance limit			8.0
NPW	0061	4575	Dibromochloromethane	10 to 100		±40% fixed acceptance limit			6.0
NPW		4570	1,2-Dibromo-3-chloropropane (DBCP)	15 to 150		±40% fixed acceptance limit			9.0
NPW		4585	1,2-Dibromoethane (EDB)	10 to 120		±35% fixed acceptance limit			6.5
NPW		4595	Dibromomethane	10 to 120		±35% fixed acceptance limit			6.5
NPW		4630	1,1-Dichloroethane	10 to 150	0.9977	0.2117	0.1227	0.0174	6.4
NPW	0054	4635	1,2 Dichloroethane	15 to 150	0.9843	1.3728	0.0912	0.4693	10.6
NPW	0246	4640	1,1-Dichloroethene	10 to 150	1.0034	0.6630	0.1447	0.0521	6.2
NPW		4645	cis-1,2-Dichloroethene	10 to 150	0.9973	0.3699	0.1095	0.0036	7.0
NPW	0247	4700	trans-1,2-Dichloroethene	10 to 120		±40% fixed acceptance limit			6.0
NPW	0248	4655	1,2-Dichloropropane	10 to 150		±30% fixed acceptance limit			7.0
NPW		4680	cis-1,3-Dichloropropene	10 to 120		±35% fixed acceptance limit			6.5
NPW	0249	4685	trans-1,3-Dichloropropene	10 to 120		±35% fixed acceptance limit			6.5
NPW	0063	4975	Methylene Chloride	10 to 120		±40% fixed acceptance limit			6.0
NPW		5105	1,1,1,2-Tetrachloroethane	15 to 150		±35% fixed acceptance limit			9.8
NPW	0250	5110	1,1,2,2-Tetrachloroethane	15 to 150		±35% fixed acceptance limit			9.8
NPW	0059	5115	Tetrachloroethene	10 to 150	0.9416	-0.5063	0.1189	0.3441	4.3
NPW	0056	5160	1,1,1-Trichloroethane	10 to 100		±40% fixed acceptance limit			6.0
NPW	0251	5165	1,1,2-Trichloroethane	15 to 150		±30% fixed acceptance limit			10.5
NPW	0057	5170	Trichloroethene	10 to 100	0.9611	0.5720	0.1077	0.2478	6.2
NPW	0252	5175	Trichlorofluoromethane	20 to 120		± 60% fixed acceptance limit			8.0
NPW		5180	1,2,3-Trichloropropane	15 to 150	0.9867	-0.4721	0.1630	0.9605	4.1
NPW	0253	5235	Vinyl chloride	20 to 120		± 60% fixed acceptance limit			8.0
			Low-Level Halocarbons ¹¹	µg/L					µg/L
NPW		4570	1,2-Dibromo-3-chloropropane (DBCP)	0.2 to 2.0	0.9542	0.0359	0.1200	0.0161	0.11
NPW		4585	1,2-Dibromoethane (EDB)	0.2 to 2.0	0.9341	0.0293	0.1090	0.0239	0.08
NPW		5180	1,2,3-Trichloropropane	0.2 to 2.0	0.9284	0.0534	0.1257	0.0117	0.13



TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017

Blue = New Analyte Magenta = Changes

Matrix	EPA Analyte Code	NELAC Analyte Code	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷
					a	b	c	d	
Volatile Petroleum Hydrocarbons									
				µg/L					µg/L
NPW		9408	Gasoline range organics (GRO) ¹³	400 to 4000	1.0683	-7.7234	0.2162	35.0439	55
Base/Neutrals ¹									
				µg/L					µg/L
NPW	0189	5500	Acenaphthene	10 to 200	0.7748	0.8506	0.1427	0.1159	4.0
NPW	0190	5505	Acenaphthylene	10 to 200	0.8029	-0.2974	0.1485	0.1111	2.9
NPW	0192	5555	Anthracene	10 to 200	0.7986	1.7870	0.1229	0.7303	3.9
NPW	0177	5575	Benzo(a)anthracene	10 to 200	0.8381	0.5699	0.1162	0.6075	3.6
NPW	0254	5670	Benzyl butyl phthalate	50 to 200	0.8496	-2.1863	0.1776	0.0752	13.4
NPW	0178	5585	Benzo(b)fluoranthene	20 to 200	0.8327	0.1531	0.1497	0.1078	7.5
NPW	0179	5600	Benzo(k)fluoranthene	20 to 200	0.8223	1.996	0.1862	1.126	7.7
NPW	0180	5590	Benzo(g,h,i)perylene	10 to 200	0.8261	1.5562	0.1556	0.0166	5.1
NPW	0255	5580	Benzo(a)pyrene	10 to 200	0.8207	-0.0550	0.1484	0.4349	2.4
NPW	0198	5660	4-Bromophenyl-phenylether	20 to 200	0.8081	3.0645	0.1325	0.8996	8.6
NPW	0195	5760	bis(2-Chloroethoxy)methane	20 to 200	0.7615	0.4890	0.1193	1.5633	3.9
NPW	0196	5765	bis(2-Chloroethyl)ether	20 to 200	0.7090	2.3607	0.1529	0.4801	5.9
NPW	0197	4659	2,2'-Oxybis(1-Chloropropane) ¹⁸	30 to 200	0.7285	1.6917	0.1303	2.9025	3.1
NPW	0256	6065	Bis(2-ethylhexyl) phthalate	20 to 200	0.8065	2.5761	0.1474	1.6124	5.0
NPW	0204	5825	4-Chlorophenyl-phenylether	20 to 200	0.7669	3.7466	0.1417	0.2303	9.9
NPW	0203	5795	2-Chloronaphthalene	20 to 200	0.7102	2.4854	0.1477	0.5079	6.3
NPW	0181	5855	Chrysene	10 to 200	0.8180	2.3274	0.1351	0.2137	5.8
NPW	0182	5895	Dibenzo(a,h)anthracene	20 to 200	0.8079	2.3890	0.1497	0.8729	6.9
NPW		5905	Dibenzofuran	30 to 200	0.7411	2.7181	0.1159	1.0735	11.3
NPW		4610	1,2-Dichlorobenzene	20 to 200	0.6365	0.7906	0.1517	2.2155	2.0
NPW		4615	1,3-Dichlorobenzene	20 to 200	0.5921	3.0260	0.1787	0.3464	3.1
NPW		4620	1,4-Dichlorobenzene	20 to 200	0.5671	3.6005	0.1640	0.4826	3.7
NPW	0208	6070	Diethyl phthalate	50 to 200	0.7492	3.3637	0.1805	2.0213	8.9
NPW	0209	6135	Dimethyl phthalate	50 to 200	0.6375	3.9631	0.2524	0.8174	11.5
NPW	0205	5925	Di-n-butyl phthalate	40 to 200	0.7797	5.1233	0.1490	0.8776	15.8
NPW	0186	6185	2,4-Dinitrotoluene	20 to 200	0.8219	0.4137	0.1183	1.7449	4.5
NPW	0210	6190	2,6-Dinitrotoluene	20 to 200	0.7999	0.4770	0.1316	0.1368	8.2
NPW	0211	6200	Di-n-octyl phthalate	30 to 200	0.8186	2.8779	0.1724	1.2382	8.2
NPW	0212	6265	Fluoranthene	30 to 200	0.8087	2.9863	0.1272	0.0642	15.6
NPW	0213	6270	Fluorene	10 to 200	0.7619	3.7583	0.1165	1.0349	4.8
NPW	0214	6275	Hexachlorobenzene	20 to 200	0.8202	0.2263	0.1238	0.1297	8.8
NPW	0215	4835	Hexachlorobutadiene	50 to 200	0.6286	2.6591	0.1616	1.9082	4.3
NPW	0216	6285	Hexachlorocyclopentadiene	50 to 200	0.6216	-4.4226	0.2049	4.3222	5.0
NPW	0217	4840	Hexachloroethane	50 to 200	0.5921	-0.0657	0.1640	0.5308	3.3
NPW	0218	6315	Indeno(1,2,3, cd)pyrene	30 to 200	0.7115	5.0289	0.1430	1.4299	9.2
NPW	0219	6320	Isophorone	20 to 200	0.7981	0.7053	0.1437	0.3000	7.1
NPW		6385	2-Methylnaphthalene	20 to 200	0.6983	2.0844	0.1361	2.1436	2.0



TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017

Blue = New Analyte Magenta = Changes

Matrix	EPA Analyte Code	NELAC Analyte Code	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷	
					a	b	c	d		
Base/Neutrals ¹ cont'					$\mu\text{g/L}$					$\mu\text{g/L}$
NPW	0222	5005	Naphthalene	20 to 200	0.6749	3.5514	0.1441	1.2975	4.5	
NPW	0226	5015	Nitrobenzene	20 to 200	0.7463	0.9864	0.1388	0.4589	6.2	
NPW	0227	6530	N-Nitrosodimethylamine	75 to 200	0.4665	7.3433	0.1652	3.9997	7.5	
NPW	0230	6545	N-Nitroso-di-n-propylamine	30 to 200	0.7913	-0.0510	0.1541	0.1328	9.4	
NPW	0229	6535	N-Nitrosodiphenylamine	30 to 200	0.7740	0.6711	0.2016	0.0494	5.6	
NPW	0231	6615	Phenanthrene	10 to 200	0.8001	2.8698	0.1110	0.9485	4.7	
NPW	0187	6665	Pyrene	10 to 200	0.8476	1.0097	0.1490	0.0530	4.9	
NPW	0092	5155	1,2,4-Trichlorobenzene	20 to 200	0.6769	1.1166	0.1493	1.8128	2.0	
Acids ¹					$\mu\text{g/L}$					$\mu\text{g/L}$
NPW	0161	5700	4-Chloro-3-methylphenol	30 to 200	0.7998	0.6264	0.1421	0.0397	11.7	
NPW	0162	5800	2-Chlorophenol	30 to 200	0.7292	1.4640	0.1518	0.0174	9.6	
NPW	0163	6000	2,4-Dichlorophenol	30 to 200	0.7362	2.8458	0.1433	0.0585	11.9	
NPW		6005	2,6-Dichlorophenol	30 to 200	0.7512	3.7563	0.1564	0.0312	12.1	
NPW	0165	6130	2,4-Dimethylphenol	40 to 200	0.7496	1.4509	0.1601	0.0953	11.9	
NPW	0167	6175	2,4-Dinitrophenol	100 to 200	0.6531	3.5920	0.1695	8.5727	10	
NPW	0168	6360	2-Methyl-4,6-Dinitrophenol	40 to 200	0.8108	3.6290	0.1573	2.1683	10.7	
NPW		6400	2-Methylphenol (o-Cresol)	40 to 200	0.6821	2.2126	0.1529	0.5485	9.5	
NPW		6410	4-Methylphenol (p-Cresol) ¹⁴	50 to 200	0.6531	2.1854	0.2008	0.7807	5.0	
NPW	0171	6490	2-Nitrophenol	50 to 200	0.7631	1.1486	0.1272	2.4547	12.9	
NPW	0173	6500	4-Nitrophenol	100 to 200	0.5591	-1.0075	0.2511	1.9409	10	
NPW	0174	6625	Phenol	100 to 200	0.557	0.5929	0.253	1.0269	10	
NPW	0158	6605	Pentachlorophenol	40 to 200	0.8469	-0.7338	0.1561	1.5178	9.9	
NPW	0175	6835	2,4,5-Trichlorophenol	30 to 200	0.7726	3.2199	0.1362	0.9916	11.2	
NPW	0159	6840	2,4,6-Trichlorophenol	30 to 200	0.7880	0.8051	0.1406	0.0280	11.7	



TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017

Blue = New Analyte Magenta = Changes

Matrix	EPA Analyte Code	NELAC Analyte Code	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷
					a	b	c	d	
PCBs in Water ²									
				µg/L					µg/L
NPW	0040	8880	Aroclor 1016	2.0 to 10	0.8318	0.1991	0.1591	0.0384	0.8
NPW	0041	8885	Aroclor 1221	2.0 to 10	0.8318	0.1991	0.1591	0.0384	0.8
NPW	0042	8890	Aroclor 1232	2.0 to 10	0.8318	0.1991	0.1591	0.0384	0.8
NPW	0040	8895	Aroclor 1242	2.0 to 10	0.8318	0.1991	0.1591	0.0384	0.8
NPW	0044	8900	Aroclor 1248	2.0 to 10	0.8318	0.1991	0.1591	0.0384	0.8
NPW	0045	8905	Aroclor 1254	2.0 to 10	0.8318	0.1991	0.1591	0.0384	0.8
NPW	0046	8910	Aroclor 1260	2.0 to 10	0.8318	0.1991	0.1591	0.0384	0.8
Organochlorine Pesticides ¹									
				µg/L					µg/L
NPW	0047	7025	Aldrin	1.0 to 15	0.8524	-0.0159	0.1655	0.0002	0.34
NPW	0079	7110	alpha-BHC	2.0 to 20	0.8996	0.0151	0.1505	0.0349	0.81
NPW	0080	7115	beta-BHC	2.0 to 20	0.8889	0.1961	0.1372	0.0777	0.92
NPW	0081	7105	delta-BHC	2.0 to 20	0.9031	0.1036	0.1525	0.0673	0.79
NPW	0082	7120	gamma-BHC (Lindane)	2.0 to 20	0.8959	0.1095	0.1528	0.0189	0.93
NPW		7240	alpha-Chlordane	1.0 to 10	0.8842	0.0542	0.1423	0.0348	0.41
NPW		7245	gamma-Chlordane	1.0 to 10	0.8617	0.1041	0.1323	0.0716	0.35
NPW	0053	7250	Chlordane (total)	3.0 to 25	0.8501	0.4121	0.1540	0.0381	1.46
NPW	0049	7355	4,4'-DDD	2.0 to 10	0.9271	0.03839	0.1227	0.1763	0.63
NPW	0050	7360	4,4'-DDE	1.0 to 10	0.8793	0.0718	0.1468	0.0395	0.39
NPW	0051	7365	4,4'-DDT	1.0 to 10	0.8987	0.1076	0.1680	0.0337	0.40
NPW	0048	7470	Dieldrin	1.0 to 15	0.9126	0.0323	0.1327	0.0240	0.47
NPW	0083	7510	Endosulfan I	4.0 to 20	0.8698	-0.0604	0.1548	0.0549	1.40
NPW	0084	7515	Endosulfan II	4.0 to 20	0.8765	0.0994	0.1490	0.0912	1.54
NPW	0085	7520	Endosulfan sulfate	4.0 to 20	0.8752	0.5312	0.1348	0.2091	1.79
NPW	0086	7540	Endrin	2.0 to 20	0.9183	0.0706	0.1594	0.0277	0.87
NPW	0087	7530	Endrin aldehyde	4.0 to 20	0.8585	0.4845	0.1571	0.2054	1.42
NPW		7535	Endrin ketone	4.0 to 20	0.8951	0.3702	0.1135	0.1902	2.0
NPW	0052	7685	Heptachlor	1.0 to 10	0.8470	0.0457	0.1596	0.0402	0.29
NPW	0078	7690	Heptachlor Epoxide (beta)	1.0 to 10	0.9176	0.0041	0.1342	0.0268	0.44
NPW	0234	7810	Methoxychlor	2.0 to 20	0.9115	0.2801	0.1467	0.2290	0.54
NPW	0241	8250	Toxaphene	20 to 100	0.8087	1.8908	0.1991	0.5080	4.59



TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017

Blue = New Analyte Magenta = Changes

Matrix	EPA Analyte Code	NELAC Analyte Code	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷
					a	b	c	d	
			Herbicides ¹	µg/L					µg/L
NPW	0257	8545	2,4-D	2 to 10	0.7204	0.2995	0.2543	0.0297	0.20
NPW	0258	8595	Dicamba	2 to 10	0.7848	0.2788	0.1754	0.1455	0.40
NPW	0140	8655	2,4,5-T	2 to 10	0.8132	0.1393	0.1850	0.1353	0.20
NPW	0259	8650	2,4,5-TP (Silvex)	2 to 10	0.8349	0.1516	0.2046	0.0195	0.50
			Low Level PAHs ¹	µg/L					µg/L
NPW		5500	Acenaphthene	2.0 to 20	0.7600	0.1476	0.1456	0.0021	0.79
NPW		5505	Acenaphthylene	2.0 to 20	0.7856	0.0418	0.1133	0.0687	0.73
NPW		5555	Anthracene	0.5 to 5.0	0.8151	0.0194	0.1714	0.0115	0.14
NPW		5575	Benzo(a)anthracene	0.5 to 5.0	0.9012	-0.0236	0.0614	0.0462	0.20
NPW		5580	Benzo(a)pyrene	0.5 to 5.0	0.7745	0.0824	0.1162	0.0270	0.21
NPW		5585	Benzo(b)fluoranthene	0.5 to 5.0	0.8217	0.0544	0.1167	0.0144	0.25
NPW		5590	Benzo(g,h,i)perylene	0.5 to 5.0	0.7683	0.0737	0.1641	0.0088	0.18
NPW		5600	Benzo(k)fluoranthene	0.5 to 5.0	0.8943	-0.0069	0.1245	0.0108	0.22
NPW		5855	Chrysene	0.5 to 5.0	0.8883	0.0132	0.1046	0.0235	0.23
NPW		5895	Dibenz(a,h)anthracene	0.5 to 5.0	0.7914	0.0640	0.1377	0.0520	0.10
NPW		6265	Fluoranthene	0.5 to 5.0	0.8565	0.0211	0.1064	0.0128	0.25
NPW		6270	Fluorene	2.0 to 10	0.7863	0.0472	0.1153	0.0631	0.74
NPW		6315	Indeno(1,2,3-cd)pyrene	0.5 to 5.0	0.8224	0.0623	0.1316	0.0267	0.20
NPW		5005	Naphthalene	2.0 to 10	0.7279	0.0977	0.1251	0.0803	0.56
NPW		6615	Phenanthrene	0.5 to 5.0	0.8332	0.0256	0.1099	0.0118	0.24
NPW		6665	Pyrene	0.5 to 5.0	0.8468	0.0435	0.1023	0.0095	0.28
			Petroleum Hydrocarbons						
NPW		9369	Diesel range organics (DRO) ¹⁵	800 to 6000 µg/L	0.7790	-96.0467	0.1386	109.1897	80 µg/L
NPW	0104	1803	n-Hexane Extractable Material (O&G) ^{10k,16}	20 to 200 mg/L	0.9400	-0.4116	0.0545	2.0789	8.8 mg/L
NPW		1853	non-Polar Extractable Material (TPH) ¹⁷	20 to 200 mg/L	0.9692	-1.1573	0.1586	0.3709	7.6 mg/L



TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017

Blue = New Analyte

Magenta = Changes

Matrix	EPA Analyte Code	NELAC Analyte Code	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷
					a	b	c	d	

1) For volatiles, base/neutrals, acids, organochlorine pesticides, herbicides, and low level PAHs, providers must include a minimum number of analytes using the criteria described below:

PT samples that are to be scored for one to ten analytes must include all of these analytes.

PT samples that are to be scored for ten to twenty analytes must include at least ten of these analytes or 80% of the total, whichever number is greater.

PT samples that are to be scored for more than twenty analytes must include at least sixteen of these analytes or 60% of the total, whichever number is greater.

If the calculated percentage of the total number of analytes in the PT sample is a fraction, the fraction shall be rounded up to the next whole number.

2) One sample (minimum) in every study, containing one Aroclor, selected at random from among the Aroclors listed above.

3) Acceptance limits are set at the Mean \pm 3 SD

Where the a, b, c and d factors are presented, Mean = $a \cdot T + b$; SD = $c \cdot T + d$ where T is the assigned value.

Where only the c and d factors are presented, Mean = Robust Study Mean; SD = $c \cdot X + d$ where X is the Robust Study Mean.

Where no factors are presented (Study Mean \pm 3SD), Mean = Robust Study Mean, SD = Robust Study Standard Deviation.

Robust Study Mean and Standard Deviation are generated using statistical analysis of study data set. (ie. Bi-weight, Grubbs, Dixon, etc.)

Quantitative Microbiology acceptance criteria are based on the robust participant Mean and SD determined from each respective PT study

4) If the lower acceptance limit generated using the criteria contained in this table is less than (<) 10% of the assigned value, the lower acceptance limits are set at 10% of the assigned value with the exception of microbiology analytes.

5) If the lower acceptance limit generated using the criteria contained in this table is greater than 90% of the assigned value, the lower acceptance limits are set at 90% of the assigned value with the exception of microbiology analytes.

6) If the upper acceptance limit generated using the criteria contained in this table is less than 110% of the assigned value, the upper acceptance limits are set at 110% of the assigned value with the exception of microbiology analytes.

7) NELAC Proficiency Testing Reporting Limits (PTRLs) are provided as guidance to laboratories analyzing NELAC PT samples. These levels are the lowest acceptable results that could be obtained from the lowest spike level for each analyte. The laboratory should report any positive result down to the PTRL.

It is recognized that in some cases (especially for analytes that typically exhibit low recovery) the PTRL may be below the standard laboratory reporting limit. However, the laboratory should use a method that is sensitive enough to generate results at the PTRL shown. NELAC PTRLs are also provided as guidance to PT Providers. At a minimum for all analytes with an assigned value equal to "0", the PT Provider should verify that the sample does not contain the analyte at a concentration greater than or equal to the PTRL.

8) These limits are for quantitative methods using membrane filtration techniques.

9) These limits are for quantitative methods using most probable number techniques.



TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017

Blue = New Analyte

Magenta = Changes

Matrix	EPA	NELAC	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷
	Analyte Code	Analyte Code			a	b	c	d	

10) The following recommended sample designs, which were used in past USEPA studies, should be used as model designs because other designs may not give equivalent statistics. PT study providers may vary their sample designs from those shown. The specifics within each sample are within the discretion of the PT study Provider.

- a) Design criterion for Mercury – 1:1 (mole:mole as Hg) Mercuric Oxide and Methyl Mercuric Chloride.
- b) Design criteria for Demands – 1:1 Glucose and Glutamic Acid.
- c) Design criteria for 5-Day BOD and Carbonaceous BOD – The assigned value used for BOD and CBOD is derived from the linear relationship between the BOD or CBOD value and the concentration of Glucose-Glutamic Acid (GGA) or Potassium Hydrogen Phthalate (KHP) used for the formulation. For example, 150 mg/L each of Glucose & of Glutamic Acid produces a BOD of 198 mg/L, and 300 mg/L KHP produces a BOD of 240 mg/L. 0 mg/L GGA or KHP would produce a BOD value of 0 mg/L.
- d) Design criterion for Chemical Oxygen Demand – The assigned value of COD is (1.066 times mg Glucose plus 0.9787 times mg Glutamic Acid) divided by total liters of sample adjusted for required dilutions.
- e) Design criterion for Total Organic Carbon – The assigned value of TOC is (0.4000 times mg Glucose plus 0.4082 times mg Glutamic Acid) divided by total liters of sample adjusted for required dilutions.
- f) Design criterion for Total Kjeldahl Nitrogen – Glycine is the source of TKN.
- g) Design criterion for pH – in separate solution (use buffer formulation from the CRC chemical handbook).
- h) Design criterion for Total Cyanide – Potassium Ferricyanide.
- i) Design criterion for Total Phenolics (4AAP) – 40% Phenol, 20% 2-Chlorophenol, 20% 2,4-Dinitrophenol, 20% 2,4-Dichlorophenol (mole %), calculated as mg/L Phenol.
- j) Design criterion for Turbidity - Formazin is the source for Turbidity.
- k) Design criterion for Oil and Grease – 1:1 Paraffin oil and cooking oil, vacuum pump oil, or similar mixture that does not contain volatile organics.

11) The Low Level Analytes' concentration ranges and acceptance criteria are specifically intended for technologies/methods that can achieve the listed PTRL.

12) Volatiles Aromatics must contain all three Xylene isomers. The concentration range of o-Xylene and m&p-Xylene is 10-150 µg/L each.

13) Gasoline Range Organics (GRO) per purge-and-trap extraction followed by chromatographic analysis. GRO is defined as the carbon range between n-C5 and n-C10.



**TNI/NELAP PT for Accreditation
Fields of Proficiency Testing with PTRLs
Non-Potable Water (NPW)
Effective: July 24, 2017**

Blue = New Analyte

Magenta = Changes

Matrix	EPA Analyte Code	NELAC Analyte Code	Analyte ^{1,2}	Conc Range	Acceptance Criteria ^{3,4,5,6}				NELAC PTRL ⁷
					a	b	c	d	
			14) Laboratories seeking or maintaining NELAP accreditation for Non-Potable Water 4-Methylphenol or the coeluting isomer pair of 3-Methylphenol and 4-Methylphenol must meet the NELAC PT requirements for this Field of Proficiency Testing (4-Methylphenol).						
			15) Diesel Range Organics (DRO) per solvent extraction followed by chromatographic analysis. DRO is defined as the carbon range between n-C ₁₀ and n-C ₂₈ .						
			16) n-Hexane Extractable Material (HEM) per solvent extraction followed by gravimetric or infrared spectrometric analysis (Oil & Grease).						
			17) non-Polar Extractable Material per solvent extraction and Silica Gel Treated (SGT) followed by gravimetric or infrared spectrometric analysis (Total Petroleum Hydrocarbons).						
			18) Also known as Bis(2-chloro-1-methylethyl) Ether, formerly known as Bis(2-chloroisopropyl) Ether.						